

## Quality Control Test Certificate

Product name	CytoSure Genomic DNA Labelling Kit
Part no.	020020
Batch no.	9842715
Certificate issue no.	1

### Acceptance Criteria

Test		Specification	Pass/fail
Dye incorporation	Cy3	>250 pmol	PASS
	Cy5	>250 pmol	PASS
DNA yield		>6µg	PASS
Array performance (Derivative Log Ratio Spread)		<0.25	PASS
Date 22 <sup>nd</sup> May 2017			
Initials DH			

This product has been manufactured and tested by Oxford Gene Technology (Operations) Ltd. under an ISO 9001:2008 certified management system.

### Oxford Gene Technology

T: +44(0)1865 856826 (US: 914-467-5285) E: products@ogt.com W: www.ogt.com

Technical support: support@ogt.com

The purchaser has the non-transferable right to use and consume the product for RESEARCH USE ONLY AND NOT for DIAGNOSTICS PROCEDURES. It is not intended for use, and should not be used for the diagnosis, prevention, monitoring, treatment or alleviation of any disease or condition, or for the investigation of any physiological process, in any identifiable human, or for any other medical purpose.