

Quality Control Test Certificate

Product name	SureSeq NGS Hyb & Wash Kit
Product no.	500075
Batch no.	067203
Certificate issue no.	1

Acceptance Criteria

The SureSeq NGS Hyb & Wash Kit was used with the short and long hyb. protocols alongside a SureSeq Core MPN and SureSeq myPanel NGS Custom Cancer Panel respectively; and then sequenced on an Illumina sequencer to verify that it meets the below series of quality specifications:

Number of Samples per MiSeq Lane: 16 DNA Input: 1000ng

		Specification	Pass/Fail
Mean Target Coverage:	Core MPN:	> 3000x	PASS
	Custom myPanel:	> 1000x	PASS
% Target Bases Covered at >20% Mean:		>99.0%	PASS

This product has been approved for release by:

Name: Vincent Wong Position: Quality Associate Date: 11 December 2019

This product has been manufactured and tested by Oxford Gene Technology (Operations) Ltd. under an ISO 9001:2015 certified management system.

Oxford Gene Technology

T: +44(0)1865 856826 (US: 914-467-5285) E: products@ogt.com W: www.ogt.com
 Technical Support: support@ogt.com

The purchaser has the non-transferable right to use and consume the product for RESEARCH USE ONLY AND NOT for DIAGNOSTICS PROCEDURES. It is not intended for use, and should not be used for the diagnosis, prevention, monitoring, treatment or alleviation of any disease or condition, or for the investigation of any physiological process, in any identifiable human, or for any other medical purpose.