

QC Certificate

| | |
|--------------|---|
| Product Name | CytoSure Genomic DNA Labelling Kit (24) |
| Product Code | 020020 |
| Lot Number | 18077312 |

Acceptance Criteria

| Test | Specification | Pass/Fail | |
|---|---------------|-----------|------|
| Dye incorporation | Cy3 | >250 pmol | Pass |
| | Cy5 | >250 pmol | Pass |
| DNA yield | >6µg | Pass | |
| Array performance (Derivative Log Ratio Spread) | <0.25 | Pass | |

| Name | Position | Date |
|---------------|------------|-------------------|
| Kaajal Reeves | QC Manager | 28 September 2023 |

This product has been released by Oxford Gene Technology (Operations) Ltd. under an ISO 9001:2015 certified management system.

OGT

T: +44(0)1865 856800 (US: 914-467-5285) E: products@ogt.com W: www.ogt.com

Technical Support: support@ogt.com

The purchaser has the non-transferable right to use and consume the product for RESEARCH USE ONLY AND NOT for DIAGNOSTICS PROCEDURES. It is not intended for use, and should not be used for the diagnosis, prevention, monitoring, treatment or alleviation of any disease or condition, or for the investigation of any physiological process, in any identifiable human, or for any other medical purpose.